

Decontamination Area

Richard Wolf Flexible Endoscopes



- Perform a Leakage Test!
- If pass, continue with cleaning process.
- If fail, stop and send for repair.



- Disassemble and soak in enzymatic detergent.



- Flush, brush, flush all channels, accessories, scope exterior.



- Rinse all channels, accessories, and scope with processed water and correct pressure device.